



Research article

Learning process for identifying different types of communication via repetitive stimulation: feasibility study in a cultured neuronal network

Yoshi Nishitani^{1,*}, Chie Hosokawa², Yuko Mizuno-Matsumoto³, Tomomitsu Miyoshi⁴ and Shinichi Tamura⁵

¹ Department of Radiology, Graduate School of Medicine, Osaka University, Suita 565-0871, Japan

² Graduate School of Science Osaka City University, Osaka, 558-8585, Japan

³ Graduate School of Applied Informatics, University of Hyogo, Kobe 650-0044, Japan

⁴ Department of Integrative Physiology, Graduate School of Medicine, Osaka University, Suita 565-0871, Japan

⁵ NBL Technovator Co., Ltd., 631 Shindachimakino, Sennan 590-0522, Japan

* **Correspondence:** E-mail: ynishitani1027@gmail.com; Tel: +819036276879.

Supplementary Figure

Figure S1 shows the deposition of electrodes in the microelectrode array, MED-P545A (Alpha MED Scientific, Kadoma, Osaka, Japan). The size of, and spacing between the electrodes was $50 \times 50 \mu\text{m}^2$ and $450 \mu\text{m}$, respectively. Each electrode corresponds to a channel of spike recording. The number indicates the electrode (channel) number.

1	2	3	4	5	6	7	8
9	10	11	12	13	14	15	16
17	18	19	20	21	22	23	24
25	26	27	28	29	30	31	32
33	34	35	36	37	38	39	40
41	42	43	44	45	46	47	48
49	50	51	52	53	54	55	56
57	58	59	60	61	62	63	64

Figure S1. Deposition of electrodes in an MED-P545A microelectrode array.



AIMS Press

© 2019 the Author(s), licensee AIMS Press. This is an open access article distributed under the terms of the Creative Commons Attribution License (<http://creativecommons.org/licenses/by/4.0>)